

DATE _____
SH. 1 OF 1
COMPILED BY _____

TMC SPECIFICATION NO. S - 467

TITLE: METER CONVERSION FOR GPT-40K

JOB _____

APPROVED

PURPOSE:

To convert meters, where possible, for use in equipments other than for which originally designed.
This method will be used only where the buying of the desired meter is time consuming and providing meters for conversion are available from TMC stock.

TO CONVERT TO MR-129

Use MR-117 Meter (Remove Dial Face)
External Shunt .00415 ohm
LD-647 Dial Face

TO CONVERT TO MR-131

Use MR-121 Meter (" ")
External Rectifier
LD-646 Dial Face

TO CONVERT TO MR-132

Use MR-120 Meter (Remove Dial Face)
LD-648 Dial Face

TO CONVERT TO MR-133

Use MR-121 Meter (" ")
LD-645 Dial Face

TO CONVERT TO MR-134

Use MR-120 Meter (" ")
LD-644 Dial Face

TO CONVERT TO MR-135

Use MR-120 Meter (" ")
LD-649 Dial Face

TO CONVERT TO MR-136

Use MR-120 Meter (" ")
LD- Dial Face

DO NOT
ISSUE

SEE M. STURMER

1-28-60